




Australian Government

Department of Sustainability, Environment,
Water, Population and Communities
Supervising Scientist

*internal
report*

576



eriss HPGe detector calibration

J Pfitzner

October 2010

(Release status – unrestricted)

This page has been left blank intentionally.

***eriss* HPGe detector calibration**

J Pfitzner

Supervising Scientist Division
GPO Box 461, Darwin NT 0801

October 2010

Registry File SSD2002/0216

(Release status – unrestricted)



Australian Government

Department of Sustainability, Environment, Water, Population and Communities
Supervising Scientist

How to cite this report:

Pfitzner J 2010. **eriss** HPGe detector calibration. Internal Report 576, October, Supervising Scientist, Darwin. Unpublished paper.

Location of final PDF file in SSDX Sharepoint:

[Supervising Scientist Division > PublicationWork > Publications and Productions > Internal Reports \(IRs\) > Nos 500 to 599 > IR576 HPGe detector calibration \(Pfitzner\)](#)

Location of all key data files for this report in SSDX Sharepoint:

[Supervising Scientist Division > SSDX > Equipment and Techniques > HPGe Gamma Spec > Calibrations > HPGe Cal IR Files](#)

Author of this report:

John Pfitzner – Environmental Research Institute of the Supervising Scientist, GPO Box 461, Darwin NT 0801, Australia.

Additional contact: Andreas Bollhöfer, Program Leader, Environmental Radioactivity, Environmental Research Institute of the Supervising Scientist, GPO Box 461, Darwin NT 0801, Australia.

The Supervising Scientist is part of the Australian Government Department of Sustainability, Environment, Water, Population and Communities.

© Commonwealth of Australia 2010

Supervising Scientist

Department of Sustainability, Environment, Water, Population and Communities

GPO Box 461, Darwin NT 0801 Australia

This work is copyright. Apart from any use as permitted under the Copyright Act 1968, no part may be reproduced by any process without prior written permission from the Supervising Scientist. Requests and inquiries concerning reproduction and rights should be addressed to Publications Inquiries, Supervising Scientist, GPO Box 461, Darwin NT 0801.

e-mail: publications_ssd@environment.gov.au

Internet: www.environment.gov.au/ssd (www.environment.gov.au/ssd/publications)

The views and opinions expressed in this report do not necessarily reflect those of the Commonwealth of Australia. While reasonable efforts have been made to ensure that the contents of this report are factually correct, some essential data rely on references cited and/or the data and/or information of other parties, and the Supervising Scientist and the Commonwealth of Australia do not accept responsibility for the accuracy, currency or completeness of the contents of this report, and shall not be liable for any loss or damage that may be occasioned directly or indirectly through the use of, or reliance on, the report. Readers should exercise their own skill and judgment with respect to their use of the material contained in this report.

Printed and bound in Darwin NT by Supervising Scientist Division

Contents

Executive summary	iv
Introduction	1
The calibration and calculation procedures	2
Calibration flow chart	2
Calculations in Visual Gamma	3
Flow chart of initial calculations in Excel spreadsheets on 'P' file data	5
Calculations in the Excel spreadsheet	7
^{238}U calibration factor standards and ^{226}Ra ratio standards	9
Flow chart of radium (^{226}Ra) ratio calculations	10
Flow chart of calculation for ^{238}U series and ^{235}U calibration	11
^{232}Th calibration factor standards and ratio standards	15
Flow chart of ^{232}Th series calibration	16
Flow chart for ^{232}Th ratios at 63keV	18
Flow chart for ^{40}K and ^{137}Cs calibration	19
Entering calibration data into Visual Gamma	21
Validation procedure	22
Other Quality Assurance Procedures	23
References	24
Appendix 1 Table of geometry designation, size, containment material, sample type and sample weight range	25
Appendix 2 Current Calibration Data	26
Appendix 3 Plots of absolute efficiencies for Detectors G, N and O	42
Appendix 4 Tables of absolute efficiency values for detectors G, N and O	50
Appendix 5 Listing of Gamma Spectroscopy Standards Used for calibration	56

Executive summary

The Environmental Radioactivity Group (EnRad) of the Supervising Scientist Division (SSD) has undertaken a comprehensive calibration of their High Purity Germanium (HPGe) gamma spectroscopy system.

Three detectors have been calibrated for five different geometries (sample types) holding sample masses of less than a gram up to 150 grams. This IR documents the calibration procedure, the methods of calculation and the results of the system calibration.

***eriss* HPGe detector calibration**

J Pfitzner

Introduction

The Environmental Radioactivity (EnRad) Group in the Supervising Scientist Division (SSD) operates several High Purity Germanium (HPGe) detectors for the determination of radionuclide activity concentration by gamma spectroscopy analysis. Samples are analysed for naturally occurring radionuclides of the uranium and thorium decay series and ^{40}K , as well as for the nuclear bomb fallout isotope ^{137}Cs . These analyses are done for internal research projects and on a consultancy basis for external clients.

For quality assurance purposes, routine background and calibration checks are conducted to ensure that the instruments are operating within specifications, and to ensure that accurate results for these analyses are reported. Replacing hardware or making adjustments to instrument settings can cause significant changes to the calibration. Recent changes to the hardware triggered the need to do the calibration described in this document. All three detectors (detectors N, G and O) in current use were calibrated.

Samples are prepared in one of several 'geometries'. Each geometry is assigned an identifying letter. The 'geometries' are designed to ensure consistency in the shape and quantity of sample presented to the detector. The samples are either confined within a plastic container (geometries 'Q' and 'R') or cast in polyester resin (geometries 'A', 'B' and 'V'). The calibration was done for all five geometries for each of the three detectors.


The range of geometries allows for optimisation of counting for a variety of sample sizes and types. Geometries 'A' and 'V' are ideal for organic material, such as dried or ashed leaf litter or mussel flesh. Geometries 'B', 'Q' and 'R' are for sediments and soils and similar inorganic materials. For a more complete description of these geometries, see Appendix 1 Table of Geometry Designations, Sizes and Sample Types.

Specific radionuclide standards are prepared for each geometry. The standards are grouped in their naturally occurring ^{238}U and ^{232}Th series. Separate standards are prepared for ^{40}K and ^{137}Cs . To reduce statistical uncertainty, the activity of the standards are such that a one day count will ensure an uncertainty of less than 1% in counting statistics alone. This level of activity also ensures that the instrument background count rate is less than 0.5% of the standard activity count rate.

Standards are prepared in batches of three, to allow measurement of systematic uncertainty. Standards are typically counted for one day, whereas backgrounds are counted for three days, usually over a weekend.

The calibration and calculation procedures

Calibration flow chart



Calculations in Visual Gamma

The program *Visual Gamma* developed in-house is used to operate the detector systems; starting and stopping counting, as well as performing the spectrum analyses. The Visual Gamma Operator's Manual (Esparon & Pfitzner 2010) describes the calculation procedures for radionuclide analysis in Appendices B: *Sources of uncertainty and propagation of uncertainty equations*, C: *Equations for the calculation of weighted average* and D: *Peak area and uncertainties by end point averaging*.

For calibration, the results of the calculation of standard and matrix and/or instrument background count rates, less continuum background, are required. At an intermediate stage of the analysis process, *Visual Gamma* produces a 'P' file which contains the desired count data, and count time in kiloseconds, for calibration calculations. Fig. 1 shows a typical 'P' file.

These 'P' files are read into an Excel spreadsheet, in which all relevant counting details are retained for Quality Assurance. Isotope line count data and associated uncertainty values are copied into the Excel sheet columns for manipulation. The copying process does not separate the 'live' count time of the standard or background from text, so this value, in ksec, must be manually entered in separate cells, identified by a label 'ksec' in an adjacent cell.

The 'P' file count data for each isotope line are :

- (1) the end point averaged counts for the peak region of interest (ROI) range of channels, with the continuum, ie the average counts of the left and right background ROI channels adjacent to the peak, subtracted.

$$A = P - \frac{n}{2} \left(\frac{B_1}{n_1} + \frac{B_2}{n_2} \right)$$

where A is the end point averaged counts, P is the total counts within the peak ROI of n channels, and B₁ and B₂ are the total counts in the adjacent background ROIs of n₁ and n₂ channels respectively.

- (2) the 1 standard deviation uncertainty associated with the above calculations is calculated using:

$$\sigma A = \sqrt{\left[P + \left(\frac{n}{2} \right)^2 \left(\frac{B_1}{n_1^2} + \frac{B_2}{n_2^2} \right) \right]}$$

Usually three standards of each radionuclide series in each geometry are counted on each detector. In some cases there is only one standard for a particular radionuclide in that geometry, eg ¹³⁷Cs in A, B and V geometries.

The Excel worksheets for all the 'P' file results for one radionuclide series (U, Th, ⁴⁰K or ¹³⁷Cs) in one geometry (A, B, V, Q or R) for one detector (N, G or O) are in one Excel file. For example Det-G-Geom-Q-Th-Cal.xls contains the calibration data for detector 'G', geometry 'Q', ²³²Th calibration. This naming convention is maintained for all detectors, geometries and radionuclides.

There are two sets of ratios calculated, the ratio of ²³²Th in ²³⁴Th at the 63 keV energy line and the ratio of ²²⁶Ra in ²³⁵U at the 186keV energy line. The 'P' files of ratio calculations are incorporated into the respective Th and U series Excel spreadsheets.

Latest calibration data are stored on SSDX Sharepoint in the following location: SSDX Equipment and Techniques > HPGe Gamma Spec > Calibrations > (year).

Spectrum File: ZV034o.01

Description : ZV034

Cal File : Cal9o.100

Analyst : John Pfitzner

MCB : 7

Start Time : 4/05/2010 1:13:42 PM

Live Time : 77388

Mass : 0.001996 kg


Sample Counts				Test Source Counts	
Radio	Energy	Peak Area	1 StdDev	Peak Area	1 StdDev
Ra-223	270	217.50	78.52	733.50	40.80
Th-234	63	-65.83	112.15	65,016.50	272.16
Pa-234	1001	-31.44	35.58	3,126.50	56.42
Th-230	68	261.86	99.44	3,180.50	105.58
Ra-226	186	11,098.73	140.27	43,098.50	214.03
Pb-214	295	35,460.16	200.23	747.00	37.81
Pb-214	352	56,213.59	248.94	906.50	37.65
Bi-214	609	33,731.99	190.56	482.00	26.79
Pb-210	46	9,463.24	132.25	37,210.00	210.01
Ac-228	338	68.00	64.59	297.50	26.86
Ac-228	911	-51.17	44.85	51.50	10.52
Ac-228	969	-31.38	38.97	23.00	8.34
Ra-224	241	16,375.51	147.57	770.50	42.23
Pb-212	239	102.31	75.74	703.00	40.74
Bi-212	40	104.35	79.36	2,281.00	87.77
Tl-208	583	72.18	48.19	175.00	20.22
K-40	1461	5.18	27.41	5.00	3.94
Cs-137*	662	-25.18	45.12	143.50	19.06
Be-7	478	-68.07	51.16	174.50	23.39

Radio	Energy	PEAK		Left Back		Right Back	
Ra-223	270	1428	1449	1418	1427	1493	1502
Th-234	63	347	360	341	346	361	367
Pa-234	1001	5328	5347	5319	5327	5372	5380
Th-230	68	371	382	364	370	383	389
Ra-226	186	1000	1016	990	999	1017	1025
Pb-214	295	1578	1596	1568	1577	1630	1638
Pb-214	352	1878	1900	1867	1877	1901	1911
Bi-214	609	3243	3267	3232	3242	3268	3279
Pb-210	46	261	273	254	260	274	279
Ac-228	338	1807	1824	1799	1806	1837	1845
Ac-228	911	4846	4871	4801	4813	4872	4883
Ac-228	969	5154	5178	5119	5131	5179	5191
Ra-224	241	1297	1313	1260	1267	1314	1322
Pb-212	239	1277	1293	1257	1264	1314	1321
Bi-212	40	224	234	217	223	235	240
Tl-208	583	3105	3127	3080	3090	3128	3138
K-40	1461	7764	7784	7753	7763	7785	7795
Cs-137*	662	3521	3544	3510	3520	3563	3574
Be-7	478	2547	2567	2535	2546	2616	2628

Figure 1 A typical 'P' file, for radium ratio standard SV034 counted on detector 'O'

Flow chart of initial calculations in Excel spreadsheets on 'P' file data

The instructions flowsheet below assumes that the operator uses the template files that have been previously prepared for the calibration process. These are located on the SSDX Sharepoint path: SSDX Equipment and Techniques > HPGe Gamma Spec > Calibrations. SSDX Sharepoint is the Division's electronic document management and retrieval system (EDRMS). A typical Excel worksheet for calculation of the calibration factor, Cts/Ksec/Bq for a ^{137}Cs standard in 'A' geometry on detector 'N' is shown in Figure 2.



Spectrum File: ZA034n.01												
Description:ZA034		Live count time, ksec			Total Activity, Bq		Peak Area/ live count time				Peak Area less Background Cts/Ksec / total activity, Bq	
Cal File : Cal1n.103												
Analyst : John Pfitzner		137Cs			Tot Bq	Mar2010						
MCB : 1												
Start Time: 20/04/2010 2:08:43 PM												
Live Time: 79513		ksec			79.513	ksec						
Mass: 13.6365 kg												
		81.27			Bq 137Cs	Mar1983						
Sample Counts							Cts/Ksec	Cts/ksec	1sd	1sduncert	Cts/Ksec/Bq	1sd
Radio	Energy	Peak Area			1StdDev	Peak	less bgd	uncert	less bgd	Peak		uncert
Ra-223	270	74.22			65.81	0.9334	0.9871	0.8277	0.8368			
Th-234	63	129.06			53.64	1.6231	1.0335	0.6746	0.6834			
Pa-234	1001	-1.64			13.02	-0.0206	0.0704	0.1637	0.1719			
Th-230	68	-9			51.64	-0.1132	-0.6902	0.6495	0.6587			
Ra-226	186	186.93			82.69	2.3509	1.6793	1.0400	1.0518			
Pb-214	295	39.55			61.36	0.4974	-0.0961	0.7717	0.7803			
Pb-214	352	46.39			64.19	0.5834	-0.6407	0.8073	0.8146			
Bi-214	609	48.92			31.06	0.6152	-0.3251	0.3906	0.4002			
Pb-210	46	-82			59.73	-1.0313	-2.0136	0.7512	0.7596			
Ac-228	338	-13.1			59.14	-0.1648	-0.3357	0.7438	0.7505			
Ac-228	911	-3.29			14.67	-0.0414	-0.1680	0.1845	0.1934			
Ac-228	969	21.45			14.84	0.2698	0.1486	0.1866	0.1952			
Ra-224	241	174.25			64.18	2.1915	1.8962	0.8072	0.8167			
Pb-212	239	279.25			64.99	3.5120	2.7238	0.8174	0.8272			
Bi-212	40	-618.56			63.65	-7.7794	-7.7587	0.8005	0.8082			
Tl-208	583	35			30.32	0.4402	0.0915	0.3813	0.3887			
K-40	1461	59.58			11.4	0.7493	0.1953	0.1434	0.1509			
Cs-137*	662	75,676.18			276.89	951.7460	951.6528	3.4823	3.4830	21.7717	137Cs	0.0797
Be-7	478	175.61			60.28	2.2086	2.1342	0.7581	0.7617			

Figure 2 Worksheet for calculation of the calibration factor, Cts/Ksec/Bq for a ¹³⁷Cs standard in 'A' geometry on detector 'N'

Calculations in the Excel spreadsheet

The calculations described below use the ‘net count rate’ of the *standard*. The net rate is the difference between the counts per kilosecond of the *standard* from the ‘P’ file (from which the continuum background has already been subtracted) and the *background*, also in units of counts per kilosecond. The *background* may be a count of either a *matrix* material (eg Flying Fox sand or dried beef) or the *instrument* background. Matrix materials must be prepared in the same geometry and mass as the standard they represent. The 1 standard deviation (1 s.d.) uncertainty associated with this net count rate is calculated using the formula from the Visual Gamma Operator’s Manual (Esparon & Pfitzner 2010) Appendix B: *Sources of uncertainty and propagation of uncertainty equations*.

$$s_y = \sqrt{a_1^2 s_{x_1}^2 + a_2^2 s_{x_2}^2}$$

where $y = a_1 x_1 - a_2 x_2$ (x_1 is uncertainty in a_1 and x_2 in a_2 and s_y is 1 s.d. uncertainty in y).

The total activity of the radionuclide in the standard is entered into a clearly labelled cell near the top of the worksheet. Particular attention must be paid to short lived isotopes, such as ^{137}Cs , where corrections for decay must be applied. Details of all **eriss** standards, including total activity, is maintained in the Excel spreadsheet ‘All **eriss** gamma standards.xls’, located at SSDX>Equipment and Techniques > HPGe Gamma Spec> Calibrations>Standards Listing.

If the standard is to be used only for ratios, the activity value is not used. In this case the next calculation is to select the specific cells for the ratio calculations. These are:

- a) For the ratio of ^{232}Th in ^{234}Th at the 63 keV energy line the ratios used are

63keV:338keV

63keV:911keV and

63keV:969keV.

- b) For the ratio of ^{226}Ra in ^{235}U at the 186keV energy line the ratios used are

186keV:295keV

186keV:352keV and

186keV:609keV.

The isotopes and energy lines for calibration of radionuclides in the U decay series are (see right):

^{238}U series	
Isotope	Line Energy (keV)
^{235}U	185.720
^{234}Th	63.300
^{234}Pa	1001*
^{230}Th	68*
^{226}Ra	186.211
^{214}Pb	295.224
^{214}Pb	351.932
^{214}Bi	609.312
^{210}Pb	46.539
^{214}Pb	241.997

(All line energies, except those marked with *, are obtained from the web site of the Nucleide Organisation, which maintains the latest updates of radionuclide data :

http://www.nucleide.org/DDEP_WG/DDEPdata.htm)

The isotopes and energy lines for calibration of radionuclides in the Th-series are:

²³² Th series	
Isotope	Line Energy (keV)
²²⁸ Ac	338.320
²²⁸ Ac	911.196
²²⁸ Ac	968.960
²²⁴ Ra	240.986
²¹² Pb	238.623
²¹² Bi	39.858
²⁰⁸ Tl	583.187

The isotopes and energy lines for calibration of radionuclides of the other routinely measured radionuclides are:

Nuclides	
Isotope	Line Energy (keV)
⁴⁰ K	1460.822
¹³⁷ Cs (^m Ba)	661.659

For each radionuclide line, the net count rate is divided by the total activity of that isotope in the standard. The result is counts per kilo second per Bq, the activity conversion factor. Uncertainties are propagated according to the formula found in the Visual Gamma Operator's Manual (Esparon & Pfitzner 2010) Appendix B. 'Sources of uncertainty and propagation of uncertainty equations',

$$\frac{s_y}{y} = \sqrt{\left(\frac{s_{x_1}}{x_1}\right)^2 + \left(\frac{s_{x_2}}{x_2}\right)^2}$$

where $y = x_1 x_2$.

A summary worksheet contains the weighted average of the activity conversion factor for several standards (b, c, and d) of the same type in the same geometry on the same detector. The formula for calculating the weighted average, W_m , is:

$$W_m = \frac{\frac{b}{\delta b^2} + \frac{c}{\delta c^2} + \frac{d}{\delta d^2}}{\frac{1}{\delta b^2} + \frac{1}{\delta c^2} + \frac{1}{\delta d^2}}$$

where the values and their associated uncertainties are $b \pm \delta b$, $c \pm \delta c$ and $d \pm \delta d$,

and the propagated uncertainty in the weighted mean is

$$UncertW_m = \frac{1}{\sqrt{\frac{1}{\delta b^2} + \frac{1}{\delta c^2} + \frac{1}{\delta d^2}}}$$

The weighted averages and uncertainties for the isotope lines are tabulated for each radionuclide line in the series. The summary page also contains the weighted averages of the ratios for that series.

The systematic uncertainty is now added to the propagated uncertainty using the formula $s_y = \sqrt{a_1^2 s_{x_1}^2 + a_2^2 s_{x_2}^2}$ from above.

These systematic uncertainties are (Marten 1992):

Uranium (^{238}U) series	1.2%
Thorium (^{232}Th) series	1.5%
Others (^{137}Cs & ^{40}K)	2%


^{238}U calibration factor standards and ^{226}Ra ratio standards

eriss uses Canmet BL-5 as the standard reference material for ^{238}U series calibration. BL-5 is a uranium ore in secular equilibrium with its progeny and the activity concentration of BL-5 is accurately and precisely known.


The source material for ^{226}Ra can be one of several ^{226}Ra solutions which must not contain any ^{238}U uranium.


Flow chart of radium (²²⁶Ra) ratio calculations

The ²²⁶Ra ratio must be completed before calculating the ²³⁸U calibration series as this ratio is used in the ²³⁵U calibration at 186keV.



Flow chart of calculation for ^{238}U series and ^{235}U calibration





	Cts/Ksec/Bq							
	Peak		uncert					
	0.739585		0.010612					
	5.459564		0.021811					
	0.152224		0.004682		From radium ratios			
	0.659348		0.01675		226Ra Contribution to 186kev			
	6.401413		0.020378		Cts/ksec		Uncert	
	11.87108		0.024049		1204.9		5.644739	
	19.21909		0.02996		1227.086		5.314179	
	11.74032		0.0232		1228.531		5.715505	
	5.481455		0.020314		Wtd Ave.			
	0.029694		0.007917		1220.392		3.204102	
	0.000369		0.005096					
	0.005775		0.004384		Net 235U at 186keV			
	5.614979		0.018243		Cts/ksec		uncert	
	0.034385		0.009511		486.2248		6.307214	
	0.049522		0.012998					
	-0.00365		0.004968		235U at 186keV			
	-0.00078		0.003335		Cts/ksec/Bq		uncert	
	0.001232		0.004804		1.823799		0.023658	
	0.006448		0.006696					
					226Ra at 186keV			
					Cts/ksec/Bq		uncert	
					4.577614		0.012018	

Figure 3 Excel worksheet showing the extra calculations required for ^{235}U calibration factor at 186keV

Det - G		Geom - Q				
	Radium Ratios:		Combined	Counting		
		wtd ave	uncert 1sd	uncert 1sd		
	186:295	0.38072	0.00161	0.00161		
	186:352	0.23949	0.00097	0.00097		
	186:609	0.39251	0.00165	0.00165		
		Using:	ZQ056g.01	ZQ055g.01	ZQ019g.01	
	Uranium Efficiencies:		wtd ave	Combined	system	counting
			Cts/ksec/Bq	uncert 1sd	uncert 1.2%	uncert 1sd
	U-235	186	41.57010	0.60936	0.49884	0.34997
	Th-234	63	5.41149	0.06668	0.06494	0.01515
	Pa-234	1001	0.15271	0.00373	0.00183	0.00324
	Th-230	68	0.62459	0.01377	0.00750	0.01155
	Ra-226	186	4.43826	0.05373	0.05326	0.00706
	Pb-214	242	5.45669	0.06668	0.06548	0.01260
	Pb-214	295	11.51723	0.13920	0.13821	0.01662
	Pb-214	352	18.68124	0.22513	0.22417	0.02074
	Bi-214	609	11.38448	0.13755	0.13661	0.01600
	Pb-210	46	5.43093	0.06669	0.06517	0.01414
		Using Stds:	ZQ049g.01	ZQ001g.06	ZQ003g.01	ZQ050g.01
	Updated:	235U @186 calib & uncert on 9 July 2010.				
	on:					
	Notes:					
	Systematic uncertainties (Marten 1992)					
	U: %	1.2				
	Th: %	2				
	all others %	1.5				


Figure 4 An Excel summary worksheet for U series isotopes on detector G, geometry Q. The highlighted cells are final values copied to the calibration page in Visual Gamma.

^{232}Th calibration factor standards and ratio standards

The standards for ^{232}Th ratio at 63keV must contain no uranium and the ^{232}Th series progeny must be in isotopic equilibrium for the calibration factors. The only source material currently available at *eriss* which contains no U is an 'Amersham 1906 thorium nitrate' solution. Prior to 2005, the activity concentration of this solution was considered accurately known and the solution was used for ^{232}Th series calibration factor standards as well as the ratio. Since 2005, due to a lack of confidence in the activity concentration, this solution has only been used to make ^{232}Th ratio standards. The current source material for ^{232}Th series calibration factor standards is the IAEA reference material RGTH-1 which contains a small amount of uranium and is therefore not suitable for the ^{232}Th ratios at 63keV.

Flow chart of ^{232}Th series calibration

This flow chart follows the initial calculations in the 'P' spreadsheet, see 'Flow Chart of Initial Calculations in Excel spreadsheets on 'P' file data' above.




Det - G		Geom - Q				
Thorium calibration						
			Wtd ave	Combined	counting	system
	Isotope	line keV	cts/ksec/Bq	uncert 1sd	uncert 1sd	err 2%
	Ac-228	338	6.30702	0.12691	0.01399	0.12614
	Ac-228	911	5.32166	0.10709	0.01182	0.10643
	Ac-228	969	3.09670	0.06259	0.00907	0.06193
	Ra-224	241	3.26282	0.06627	0.01154	0.06526
	Pb-212	239	34.71357	0.69493	0.03029	0.69427
	Bi-212	40	1.22936	0.02659	0.01011	0.02459
	Tl-208	583	7.99019	0.16050	0.01497	0.15980
		Using:	ZQ008g.01&2	ZQ005g.01&2	ZQ054g.01	ZQ065g.01
	Ratios		Combined	counting		
		Wtd ave	uncert 1sd	uncert 1sd		
	63:338	0.05677	0.00192	0.00192		
	63:911	0.06732	0.00228	0.00228		
	63:969	0.11558	0.00392	0.00392		
			Using:	ZQ008g.01&2	ZQ005g.01&2	
	Updated:					
	on:					
	Notes:					
	Systematic uncertainties (Marten 1992)					
	U: %	1.2				
	Th: %	2				
	all others %	1.5				

Figure 5 An Excel summary worksheet for ^{232}Th series isotopes on detector G, geometry Q. The highlighted cells are final values copied to the calibration page in Visual Gamma.


Flow chart for ^{232}Th ratios at 63keV

This flow chart follows the initial calculations in the 'P' spreadsheet, see 'Flow Chart of Initial Calculations in Excel spreadsheets on 'P' file data' above.



Flow chart for ^{40}K and ^{137}Cs calibration

Typically separate standards are made for ^{40}K and ^{137}Cs . There are several choices of source material for ^{40}K as it is naturally occurring in potassium at the isotopic concentration of 0.001167%. Analytical grade potassium salts, such as potassium sulphate and potassium nitrate are suitable sources. ^{137}Cs , on the other hand, is a product of the nuclear testing during the period 1950 – 1990 and has been deposited in varying concentrations all over the earth's terrestrial and aquatic surface. For ^{137}Cs a source solution with a certified activity concentration must be obtained from a reputable supplier, such as United Nuclear, GE Lifesciences or Oak Ridge National Lab. Nuclear Science & Technology Division.




Det - G		Geom - Q				
40K & 137Cs calibration						
			wtd ave	combined	counting	System
	Isotope	line keV	cts/ksec/Bq	uncert 1sd	uncert 1sd	err 1.5%
	K-40	1461	1.38204	0.03483	0.02799	0.0207306
	Cs-137	662	17.88292	0.28395	0.09314	0.2682438
			Using :	ZQ058g.02	ZQ057g.01	ZQ052g.01
	Entered:					
	on:					
	Notes:	included systematic uncertainties from Marten 1992				
	Systematic uncertainties (Marten 1992)					
	U: %	1.2				
	Th: %	2				
	all others %	1.5				

Figure 6 Excel spreadsheet of Detector G, Geometry Q, summary sheet for ⁴⁰K and ¹³⁷Cs calibration


Entering calibration data into Visual Gamma

The following describes the step-by-step process of entering the calibration data into *Visual Gamma*. Copies of these calibration data pages are in Appendix 1. *Current Calibration Data*.



Validation procedure

The validation procedure is an assurance check that the appropriate spectrum files have been used and calculations and data entry have been done correctly.



Other quality assurance procedures

The International Atomic Energy Agency conducts intercomparison exercises where laboratories are invited to participate in the radionuclide determination of specially prepared samples of environmental water, sediments or biota, with the aim of assessing the ability of each laboratory to accurately measure specific radionuclide activity concentrations. **eriss** have participated in several of these exercises during the past 20 years and have always achieved excellent accuracy and precision with their reported results. Following these intercomparison exercises, the sample material, with certified radionuclide concentrations, is made available for purchase as reference materials. Several of these reference materials are routinely analysed on the HPGe detectors to provide independent quality assurance of the **eriss** gamma spectroscopy calibrations and analytical procedures. The result of several years of counting these reference materials will be the subject of a separate Internal Report.

References

- Esparon A & Pfitzner J (in press) Visual Gamma – Gamma Analysis Manual. Internal Report 539, Supervising Scientist, Darwin.
- Canberra Industries Product Catalogue, Edition twelve, p25.
- Marten R 1992. Procedures for routine analysis of naturally occurring radionuclides in environmental samples by gamma-ray spectrometry with HPGe detectors. Internal report 76, Supervising Scientist for the Alligator Rivers Region, Canberra. Unpublished paper.

Appendix 1 Table of geometry designation, size, containment material, sample type and sample weight range

Geometry Designation	Size height (mm) x radius (mm)	Containment material	Sample type	Sample weight range (g)
A	10 x 65	polyester resin	organic	0 – 30
B	10 x 65	polyester resin	mineral	10 – 30
Q	4 x 45	acrylic	mineral	5 – 7
R	20 x 75	acrylic	mineral	120 – 180
V	5 x 40	polyester resin	organic	0 – 3

Appendix 2 Current calibration data

27

enradenrad

28

Close

Detector G, Geometry Q

29

EfficiencyBackGround

Detector

G

Geometry

Q

Current Setting

Cal9g.105

Date

10/06/2010 2:38:41 PM

	Efficiency (ct/ksec/Bq) (abs error 2s)		Background (ct/ksec/Bq) (abs error 2s)		ROI's (Left Background) (Peak Counts)		(Right Background)				
U-235 (186)	1.91343	0.0561	0.378125	0.18174042	Refer Ra-226				☐ Hold ROI		
Th-234 (63)	5.41149	0.13336	0.15	0.14019034	335	340	341	351		352	355
Th-234 (92.5)											
Pa-234 (1001)	0.15271	0.00745	0.059375	0.04552925	5304	5312	5313	5330	5331	5338	
Th-230 (67)	0.62459	0.02754	0.0225	0.04106904	358	363	364	374	375	380	
Ra-226 (186)	4.43826	0.10745	0.378125	0.18174042	980	988	989	1004	1005	1012	
Pb-214 (242)	5.45669	0.13336	0	0	Refer Ra-224						
Pb-214 (295)	11.51723	0.27841	0.52375	0.14705441	1561	1568	1569	1584	1585	1593	
Pb-214 (352)	18.68124	0.45026	1.1125	0.17733206	1847	1857	1867	1886	1887	1896	
Bi-214 (609)	11.38448	0.27509	0.8225	0.15132745	3224	3233	3234	3253	3254	3263	
Pb-210 (46)	5.43093	0.13337	0.394375	0.15620365	246	251	252	262	263	268	
Ac-228 (338)	6.30702	0.25383	0.3925	0.12761922	1788	1795	1796	1814	1827	1836	
Ac-228 (911)	5.32166	0.21417	0.40125	0.09790301	4788	4800	4833	4859	4860	4872	
Ac-228 (969)	3.0967	0.12519	0.18375	0.07041543	5108	5119	5141	5165	5166	5177	
Ra-224 (241)	3.26282	0.13254	0	0	1249	1256	1285	1300	1301	1308	
Pb-212 (239)	34.71357	1.38986	2.0675	0.35324212	1249	1256	1268	1284	1301	1308	
Pb-212 (40)	1.22936	0.05317	0.048125	0.06891722	211	216	217	227	228	233	
Tl-208 (583)	7.99019	0.32101	0.659375	0.12200922	3074	3083	3095	3114	3115	3124	
K-40 (1461)	1.38204	0.03483	0.425625	0.05726764	7731	7742	7743	7767	7768	7778	
Cs-137 (662)	17.88292	0.28395	0.058125	0.07747845	3498	3509	3510	3531	3550	3560	
Be-7 (478)	4.3708	0.021	0.078125	0.07268367	2516	2529	2530	2556	2616	2628	
Mn-54 (835)	0	0	0.051875	0.09607072	0	0	0	0	0	0	
Ac-228 (63/338)	0.05677	0.00384	Ra226 (186/295)	0.38072	0.00322						
Ac-228 (63/911)	0.06732	0.00456	Ra226 (186/352)	0.23949	0.00194						
Ac-228 (63/969)	0.11558	0.00784	Ra226 (186/609)	0.39251	0.00331						

Close

Password

enradenrad

Comments

Cal9g.103 - Updated 13 April 2010 (JP). Calib - OK.
Cal9g.104 - Corrected 242keV and updated 40K and 137Cs calibrations (JP), 21 Apr 2010. Calib - OK.
Cal9g.105. Updated rois from Q geometry, June 2010 (JP). Calib OK.

Detector G, Geometry R

EfficiencyBackGround

Detector

G

Geometry

R

Current Setting

Cal6g.103

Date

10/06/2010 2:36:07 PM

Hold ROI

	Efficiency (ct/ksec/Bq) (abs error 2s)		Background (ct/ksec/Bq) (abs error 2s)		ROI's (Left Background) (Peak Counts)		(Right Background)			
U-235 (186)	1.19512	0.03346	0.378125	0.18174042	Refer Ra-226					
Th-234 (63)	1.95761	0.04996	0.15	0.14019034	335	340	341	351	352	355
Th-234 (92.5)										
Pa-234 (1001)	0.08164	0.00415	0.059375	0.04552925	5304	5312	5313	5330	5331	5338
Th-230 (67)	0.2231	0.01353	0.0225	0.04106904	358	363	364	374	375	380
Ra-226 (186)	1.65728	0.04017	0.378125	0.18174042	980	988	989	1004	1005	1012
Pb-214 (242)	2.5335	0.06238	0	0	Refer Ra-224					
Pb-214 (295)	5.4344	0.13166	0.52375	0.14705441	1561	1568	1569	1584	1585	1593
Pb-214 (352)	8.98022	0.21669	1.1125	0.17733206	1847	1857	1867	1886	1887	1896
Bi-214 (609)	6.03256	0.1459	0.8225	0.15132745	3224	3233	3234	3253	3254	3263
Pb-210 (46)	1.72352	0.04364	0.394375	0.15620365	246	251	252	262	263	268
Ac-228 (338)	2.77631	0.1119	0.3925	0.12761922	1788	1795	1796	1814	1827	1836
Ac-228 (911)	2.58692	0.10415	0.40125	0.09790301	4788	4800	4833	4859	4860	4872
Ac-228 (969)	1.49994	0.06069	0.18375	0.07041543	5108	5119	5141	5165	5166	5177
Ra-224 (241)	1.3241	0.05426	0	0	1249	1256	1285	1300	1301	1308
Pb-212 (239)	14.06817	0.56344	2.0675	0.35324212	1249	1256	1268	1284	1301	1308
Pb-212 (40)	0.28988	0.01492	0.048125	0.06891722	211	216	217	227	228	233
Tl-208 (583)	3.74721	0.15061	0.659375	0.12200922	3074	3083	3095	3114	3115	3124
K-40 (1461)	0.71714	0.0343	0.425625	0.05726764	7731	7742	7743	7767	7768	7778
Cs-137 (662)	11.21323	0.33825	0.058125	0.07747845	3498	3509	3510	3531	3550	3560
Be-7 (478)	2.2171	0.2212	0.078125	0.07268367	2516	2529	2530	2556	2616	2628
Mn-54 (835)	0	0	0.051875	0.09607072	0	0	0	0	0	0
Ac-228 (63/338)	0.03744	0.00533	Ra226 (186/295)	0.30345	0.00267					
Ac-228 (63/911)	0.04015	0.00572	Ra226 (186/352)	0.18681	0.0016					
Ac-228 (63/969)	0.06922	0.00986	Ra226 (186/609)	0.27372	0.00238					

Close

Password

enradenrad

Comments

Updated by AE to include Ra ratios. JP to update later.
Cal6g.102 - Updated 13 April (JP) Cal - OK.
Cal6g.103 Updated rois from Q geometry. June 2010 (JP). Calib OK.

30

Detector G, Geometry V

31

EfficiencyBackGround

Detector

G

Geometry

V

Current Setting

Cal8g.101

Date

11/06/2010 8:20:27 AM

Efficiency

(ct/ksec/Bq) (abs error 2s)

Background

(ct/ksec/Bq) (abs error 2s)

ROI's

(Left Background) (Peak Counts) (Right Background)

U-235 (186)	3.78934	0.10345	0.378125	0.18174042	Refer Ra-226						☐ Hold ROI	
Th-234 (63)	7.67894	0.18877	0.15	0.14019034	335	340	341	351	352	355		
Th-234 (92.5)												
Pa-234 (1001)	0.20426	0.01105	0.059375	0.04552925	5304	5312	5313	5330	5331	5338		
Th-230 (67)	0.91851	0.03673	0.0225	0.04106904	358	363	364	374	375	380		
Ra-226 (186)	4.96288	0.12211	0.378125	0.18174042	980	988	989	1004	1005	1012		
Pb-214 (242)	7.63559	0.18665	0	0	Refer Ra-224						Close	
Pb-214 (295)	16.14821	0.39044	0.52375	0.14705441	1561	1568	1569	1584	1585	1593		
Pb-214 (352)	26.04896	0.62796	1.1125	0.17733205	1847	1857	1867	1886	1887	1896		
Bi-214 (609)	14.8582	0.35934	0.8225	0.15132745	3224	3233	3234	3253	3254	3263		
Pb-210 (46)	8.08191	0.19809	0.394375	0.15620365	246	251	252	262	263	268		
Ac-228 (338)	6.99691	0.281	0.3925	0.12761922	1788	1795	1796	1814	1827	1836		
Ac-228 (911)	5.87048	0.23578	0.40125	0.09790301	4788	4800	4833	4859	4860	4872		
Ac-228 (969)	3.3954	0.13681	0.18375	0.07041543	5108	5119	5141	5165	5166	5177		
Ra-224 (241)	4.17421	0.16835	0	0	1249	1256	1285	1300	1301	1308		
Pb-212 (239)	41.25052	1.65096	2.0675	0.35324212	1249	1256	1268	1284	1301	1308		
Pb-212 (40)	0.5535	0.02766	0.048125	0.06891722	211	216	217	227	228	233		
Tl-208 (583)	8.9327	0.35832	0.659375	0.12200922	3074	3083	3095	3114	3115	3124		
K-40 (1461)	1.71061	0.06029	0.425625	0.05726764	7731	7742	7743	7767	7768	7778		
Cs-137 (662)	25.56914	4.21482	0.058125	0.07747845	3498	3509	3510	3531	3550	3560		
Be-7 (478)	4.5701	0.78589	0.078125	0.07268367	2516	2529	2530	2556	2616	2628		
Mn-54 (835)	22.42	5.73	0.051875	0.09607072	0	0	0	0	0	0		
Ac-228 (63/338)	0.07031	0.01269	Ra226 (186/295)	0.30913	0.0048							
Ac-228 (63/911)	0.08498	0.01535	Ra226 (186/352)	0.1941	0.00291							
Ac-228 (63/969)	0.14657	0.02657	Ra226 (186/609)	0.32813	0.00511							

Comments

Updated by AE to include Ra ratios. JP to update later.
Cal8g.100 - Updated calibrations for all lines, except 7Be and 54Mn, 21 May 2010 (JP). Calib - OK
Cal8g.101 - Updated rois with geom Q - 11 June 2010 (JP). Calib - OK

Password

enradenrad

32

Password
enradenrad

Detector N, Geometry B

EfficiencyBackground

Detector

N

Geometry

B

Current Setting

Cal2n.103

Efficiency

(ct/ksec/Bq) (abs error 2s)

Background

(ct/ksec/Bq) (abs error 2s)

ROI's

(Left Background) (Peak Counts) (Right Background)

Date

21/05/2010 3:19:24 PM

U-235 (186)	1.93799	0.09505	0.6574418E	0.4217485E	Refer Ra-226						☐ Hold ROI
Th-234 (63)	3.82876	0.09572	0.5702325E	0.2562967E	337	343	344	356	357	362	
Th-234 (92.5)											
Pa-234 (1001)	0.13973	0.00678	0.0586046E	0.0518505E	5290	5300	5301	5320	5343	5353	Close
Th-230 (67)	0.44257	0.0209	0.0423255E	0.0698923E	362	367	368	377	378	383	
Ra-226 (186)	2.88046	0.10503	0.6574418E	0.4217485E	980	989	990	1007	1008	1016	
Pb-214 (242)	4.4149	0.1083	0	0	Refer Ra-224						
Pb-214 (295)	9.51431	0.23031	0.6223255E	0.5804466E	1558	1566	1567	1586	1587	1595	
Pb-214 (352)	15.72847	0.37932	1.1723255E	0.9883508E	1855	1865	1866	1886	1887	1895	
Bi-214 (609)	10.36468	0.25052	0.8995348E	0.6235839E	3216	3226	3227	3249	3250	3261	
Pb-210 (46)	3.85218	0.09542	1.1446511E	0.3845303E	247	253	254	267	268	273	
Ac-228 (338)	5.00186	0.20179	0.2190697E	0.3864170E	1786	1794	1795	1812	1823	1832	
Ac-228 (911)	4.6213	0.18631	0.1890697E	0.3432256E	4777	4790	4822	4848	4849	4861	
Ac-228 (969)	2.71543	0.11009	0.0946511E	0.1772991E	5066	5076	5128	5152	5153	5166	
Ra-224 (241)	2.34054	0.09616	0	0	1249	1256	1287	1303	1304	1312	
Pb-212 (239)	27.61741	1.10617	1.0616279E	2.4490191E	1249	1256	1270	1286	1304	1312	
Pb-212 (40)	0.83675	0.03813	0.0493023E	0.1124276E	216	221	222	232	233	238	
Tl-208 (583)	7.1974	0.28943	0.4048837E	0.6311266E	3066	3076	3090	3111	3112	3122	
K-40 (1461)	1.30625	0.04055	0.4044186E	0.1326345E	7708	7720	7721	7746	7747	7759	
Cs-137 (662)	21.01024	0.65373	0.0609302E	0.0756185E	3493	3503	3504	3526	3544	3553	
Be-7 (478)	3.1895	0.0945	0.0679069E	0.0765804E	2524	2531	2532	2548	2599	2607	
Mn-54 (835)	17.086	0.0915	0.1604651E	0.1282659E	4380	4427	4428	4475	4574	4621	
Ac-228 (63/338)	0.04694	0.00448	Ra226 (186/295)	0.30294	0.01461						
Ac-228 (63/911)	0.0508	0.00486	Ra226 (186/352)	0.18366	0.00859						
Ac-228 (63/969)	0.08647	0.00826	Ra226 (186/609)	0.27693	0.01318						

Comments

Updated by AE to include Ra ratios. JP updated all U & Th series ratios and efficiencies, 27-Aug-2009. Version .101a corrected error in entry of 40keV efficiency.
Cal2n.102 - Updated all but 40K, 14 April 2010 (JP). Calib - OK
Cal2n.103 - Udated 40K, 21 May 2010 (JP). Calib - OK

Password

enradenrad

34

enradenrad

35

Password
enradenrad

36

Password
enradenrad

37

enradenrad

Detector O, Geometry B

88

EfficiencyBackGround

Detector

O

Geometry

B

Current Setting

Cal2o.105

Date

15/04/2010 3:34:54 PM

	Efficiency (ct/ksec/Bq) (abs error 2s)		Background (ct/ksec/Bq) (abs error 2s)		ROI's (Left Background) (Peak Counts) (Right Background)	
U-235 (186)	2.11797	0.07516	0.4192307E	0.2141178E	Refer Ra-226	
Th-234 (63)	4.04551	0.10261	0.3115384E	0.1781427E	344	349
Th-234 (92.5)					350	363
Pa-234 (1001)	0.13498	0.0078	0.0446153E	0.06079414	5321	5329
Th-230 (67)	0.45872	0.02468	0.2898076E	0.21425191	367	373
Ra-226 (186)	2.96256	0.08438	0.4192307E	0.2141178E	989	998
Pb-214 (242)	4.54832	0.11267	0	0	5330	5347
Pb-214 (295)	9.71026	0.23571	0.4638461E	0.16903594	5372	5380
Pb-214 (352)	15.87382	0.38349	0.8123076E	0.1918088E	364	370
Bi-214 (609)	10.0566	0.24379	0.54442307	0.15384501	5321	5329
Pb-210 (46)	4.14953	0.10386	0.5623076E	0.1932043C	367	373
Ac-228 (338)	4.97034	0.20247	0.1282692E	0.1363634E	989	998
Ac-228 (911)	4.37372	0.17799	0.1123076E	0.0779042E	5330	5347
Ac-228 (969)	2.56764	0.1058	0.03211153E	0.0563718E	5372	5380
Ra-224 (241)	2.53242	0.10677	0	0	364	370
Pb-212 (239)	28.287	1.13455	0.6830769E	0.2048771E	5321	5329
Pb-212 (40)	0.86376	0.04396	0.09192307	0.12621237	367	373
Tl-208 (583)	6.88421	0.27865	0.2138461E	0.1142950C	989	998
K-40 (1461)	1.09541	0.01769	0.3532692E	0.07843371	5330	5347
Cs-137 (662)	21.36414	0.66115	0.0465384E	0.06807967	5372	5380
Be-7 (478)	0	0	0.09538461	0.1107100E	364	370
Mn-54 (835)	0	0	0.07038461	0.08622727	5321	5329
Ac-228 (63/338)	0.04825	0.00712	Ra226 (186/295)	0.30281	0.01142	
Ac-228 (63/911)	0.05483	0.00809	Ra226 (186/352)	0.1872	0.00686	
Ac-228 (63/969)	0.09339	0.0138	Ra226 (186/609)	0.29537	0.01107	

☐ Hold ROI

Close

Password

enradenrad

Comments

Updated 137Cs, 12 April 2010 JP, validated AFB. Calib - OK
Cal2o.104 - Updated 15 April 2010 (JP). Calib - OK
Cal2o.105 - Updated 40K line, 15 April 2010 (JP). Calib - OK

30

enradenrad

40


Password
enradenrad

41


Password
enradenrad

Appendix 3 Plots of absolute efficiencies for Detectors G, N and O


Detector G, Geometry A




Detector G, Geometry B




Detector G, Geometry Q




Detector G, Geometry R




Detector G, Geometry V




Detector N, Geometry A




Detector N, Geometry B




Detector N, Geometry Q




Detector N, Geometry R




Detector N, Geometry V




Detector O, Geometry A




Detector O, Geometry B




Detector O, Geometry Q



Detector O, Geometry R



Detector O, Geometry V



Appendix 4 Tables of absolute efficiency values for detectors G, N and O

Detector G, Geometry A, B & Q

	Det - G		Geom - A	Geom - A	Geom - B	Geom - B	Geom - Q	Geom - Q
	Isotope	keV	absolute efficiency %	Efficiency uncert 1sd	absolute efficiency %	Efficiency uncert 1sd	absolute efficiency %	Efficiency uncert 1sd
50	U-235	186	0.36906	0.00610	0.35139	0.00834	0.33569	0.00492
	Th-234	63	9.78298	0.13031	9.99460	0.20214	14.43063	0.17782
	Pa-234	1001	1.37543	0.05980	1.32921	0.03790	1.64507	0.04013
	Th-230	68	11.44211	0.45664	11.83883	0.34333	16.65582	0.36725
	Ra-226	186	8.05946	0.10086	7.90786	0.17021	12.48457	0.15113
	Pb-214	242	6.14021	0.07842	5.97987	0.12032	7.50783	0.09175
	Pb-214	295	5.11412	0.06293	5.03912	0.10102	6.25460	0.07560
	Pb-214	352	4.30437	0.05243	4.25408	0.08520	5.24754	0.06324
	Bi-214	609	2.10991	0.02591	2.09796	0.04205	2.50263	0.03024
	Pb-210	46	8.68988	0.11422	8.69755	0.17563	12.77264	0.15684
	Ac-228	338	4.17079	0.08384	4.12513	0.08337	5.53247	0.11133
	Ac-228	911	1.57248	0.03160	1.58180	0.03195	2.03117	0.04087
	Ac-228	969	1.50913	0.03044	1.52365	0.03099	1.94761	0.03937
	Ra-224	241	5.81515	0.11787	5.64790	0.11625	7.91946	0.16085
	Pb-212	239	6.20771	0.12425	6.15323	0.12325	7.96183	0.15939
	Bi-212	40	8.02577	0.17339	7.71466	0.18172	12.17189	0.26322
	Tl-208	583	2.10298	0.04220	2.13512	0.04299	2.61973	0.05262
	K-40	1461			1.06810	0.01682	1.30999	0.01651
	Cs-137	662	2.58279	0.04005	2.22696	0.03445	2.10387	0.01670

Detector G, Geometry R & V

Det - G		Geom - R	Geom - R	Geom - V	Geom - V
Isotope	keV	absolute efficiency %	Efficiency uncert 1sd	absolute efficiency %	Efficiency uncert 1sd
U-235	186	0.20967	0.00294	0.66480	0.00907
Th-234	63	5.22030	0.06661	20.47718	0.25169
Pa-234	1001	0.87941	0.02237	2.20036	0.05949
Th-230	68	5.94932	0.18047	24.49370	0.48972
Ra-226	186	4.66184	0.05649	13.96028	0.17175
Pb-214	242	3.48583	0.04292	10.50576	0.12840
Pb-214	295	2.95124	0.03575	8.76953	0.10602
Pb-214	352	2.52253	0.03043	7.31712	0.08820
Bi-214	609	1.32613	0.01604	3.26626	0.03950
Pb-210	46	4.05344	0.05132	19.00732	0.23294
Ac-228	338	2.43536	0.04908	6.13764	0.12325
Ac-228	911	0.98738	0.01988	2.24064	0.04500
Ac-228	969	0.94336	0.01908	2.13547	0.04302
Ra-224	241	3.21384	0.06585	10.13159	0.20430
Pb-212	239	3.22664	0.06461	9.46113	0.18933
Bi-212	40	2.87008	0.07388	5.48017	0.13693
Tl-208	583	1.22860	0.02469	2.92875	0.05874
K-40	1461	0.67975	0.01626	1.62143	0.02858
Cs-137	662	1.31920	0.01990	3.00813	0.24793

Detector N, Geometry A, B & Q

52

Det - N		Geom - A		Geom - B		Geom - Q	
		absolute	Efficiency	absolute	Efficiency	absolute	Efficiency
Isotope	keV	efficiency %	uncert 1sd	efficiency %	uncert 1sd	efficiency %	uncert 1sd
U-235	186	0.3575	0.0073	0.3400	0.0083	0.2915	0.0063
Th-234	63	10.8583	0.1448	10.2100	0.1276	12.7357	0.1627
Pa-234	1001	1.4469	0.0618	1.5052	0.0365	1.6456	0.0521
Th-230	68	11.7898	0.4321	11.8018	0.2787	15.1126	0.4520
Ra-226	186	8.2221	0.1204	8.1026	0.1477	11.6879	0.1500
Pb-214	242	6.2165	0.0796	6.0744	0.0745	7.2126	0.0897
Pb-214	295	5.3284	0.0656	5.1669	0.0625	6.0920	0.0742
Pb-214	352	4.5118	0.0550	4.4181	0.0533	5.1548	0.0624
Bi-214	609	2.3276	0.0286	2.2785	0.0275	2.6044	0.0316
Pb-210	46	10.3626	0.1354	9.0597	0.1122	11.5599	0.1458
Ac-228	338	4.3629	0.0880	4.3876	0.0885	5.3432	0.1079
Ac-228	911	1.7502	0.0353	1.7639	0.0356	2.1090	0.0426
Ac-228	969	1.6879	0.0342	1.7078	0.0346	2.0466	0.0416
Ra-224	241	6.0466	0.1239	5.6809	0.1167	8.8228	0.1800
Pb-212	239	6.3538	0.1272	6.3343	0.1269	7.5157	0.1505
Bi-212	40	9.2660	0.2083	8.2847	0.1888	11.3204	0.2580
Tl-208	583	2.3238	0.0467	2.3598	0.0474	2.7121	0.0546
K-40	1461			1.2382	0.0192	1.5427	0.0357
Cs-137	662	2.5614	0.0395	2.4718	0.0385	2.4712	0.0381

Detector N, Geometry R & V

53

Det - N		Geom - R		Geom - V	
		absolute	Efficiency	absolute	Efficiency
Isotope	keV	efficiency %	uncert 1sd	efficiency %	uncert 1sd
U-235	186	0.2026	0.0028	0.4774	0.0095
Th-234	63	5.8205	0.0725	15.3820	0.1929
Pa-234	1001	1.0523	0.0208	1.9048	0.0583
Th-230	68	6.5913	0.1614	17.3315	0.4338
Ra-226	186	5.3877	0.0652	11.2248	0.1407
Pb-214	242	3.8614	0.0470	8.3916	0.1037
Pb-214	295	3.2927	0.0397	7.1142	0.0864
Pb-214	352	2.8386	0.0342	6.0154	0.0727
Bi-214	609	1.5507	0.0187	2.9524	0.0358
Pb-210	46	4.6461	0.0576	14.7993	0.1836
Ac-228	338	2.9143	0.0584	5.3365	0.1076
Ac-228	911	1.2494	0.0250	2.1045	0.0424
Ac-228	969	1.2003	0.0241	2.0206	0.0410
Ra-224	241	3.6442	0.0732	8.4894	0.1730
Pb-212	239	3.6704	0.0734	7.8744	0.1577
Bi-212	40	3.3691	0.0710	4.2351	0.1197
Tl-208	583	1.5010	0.0300	2.7680	0.0557
K-40	1461	0.8259	0.0130	1.5145	0.0572
Cs-137	662	1.5570	0.0234	3.6620	0.4834

Detector O, Geometry A, B & Q

54

Det - O		Geom - A		Geom - B		Geom - Q	
		absolute	Efficiency	absolute	Efficiency	absolute	Efficiency
Isotope	keV	efficiency %	uncert 1sd	efficiency %	uncert 1sd	efficiency %	uncert 1sd
U-235	186	0.38834	0.00655	0.37157	0.00659	0.32910	0.00512
Th-234	63	11.87757	0.15706	10.78802	0.13681	14.71099	0.18274
Pa-234	1001	1.56469	0.06422	1.45407	0.04202	1.67273	0.04250
Th-230	68	13.36576	0.48257	12.23260	0.32910	16.05772	0.39140
Ra-226	186	8.93643	0.11185	8.33350	0.11868	12.71301	0.15474
Pb-214	242	6.58518	0.08389	6.25801	0.07751	7.64617	0.09372
Pb-214	295	5.60044	0.06876	5.27330	0.06400	6.41787	0.07766
Pb-214	352	4.71732	0.05738	4.45894	0.05386	5.38912	0.06499
Bi-214	609	2.31336	0.02835	2.21073	0.02680	2.59621	0.03140
Pb-210	46	11.50083	0.14795	9.75901	0.12213	13.19971	0.00413
Ac-228	338	4.41603	0.08872	4.35995	0.08880	5.69479	0.11463
Ac-228	911	1.69113	0.03397	1.66936	0.03397	2.14976	0.04327
Ac-228	969	1.65326	0.03331	1.61487	0.03327	2.07190	0.04189
Ra-224	241	6.66358	0.13482	6.14666	0.12958	9.18799	0.18630
Pb-212	239	6.72880	0.13466	6.48784	0.13011	8.35880	0.16734
Bi-212	40	10.19262	0.21502	8.55209	0.21763	13.17269	0.28598
Tl-208	583	2.25974	0.04533	2.25712	0.04568	2.79084	0.05607
K-40	1461			1.03830	0.00838	1.45732	0.02377
Cs-137	662	2.60320	0.04010	2.51343	0.03889	2.92420	0.04487

Detector O, Geometry R & V

55

Det - O		Geom - R		Geom - V	
		absolute	Efficiency	absolute	Efficiency
Isotope	keV	efficiency %	uncert 1sd	efficiency %	uncert 1sd
U-235	186	0.21894	0.00304	0.58543	0.01088
Th-234	63	6.23251	0.07867	19.11985	0.23503
Pa-234	1001	0.97483	0.02181	2.07615	0.05417
Th-230	68	7.06288	0.20657	21.21962	0.40013
Ra-226	186	5.61996	0.06803	13.58757	0.16909
Pb-214	242	3.99134	0.04888	9.85690	0.12032
Pb-214	295	3.37878	0.04083	8.32324	0.10054
Pb-214	352	2.88666	0.03478	6.96880	0.08396
Bi-214	609	1.51946	0.01834	3.18056	0.03842
Pb-210	46	4.94884	0.06192	18.30047	0.22367
Ac-228	338	2.92505	0.05862	5.74651	0.11546
Ac-228	911	1.19556	0.02395	2.14722	0.04314
Ac-228	969	1.14315	0.02293	2.07421	0.04182
Ra-224	241	4.11451	0.08281	9.84828	0.19890
Pb-212	239	3.74430	0.07491	8.90858	0.17829
Bi-212	40	3.60747	0.07762	5.15345	0.12805
Tl-208	583	1.45752	0.02919	2.83315	0.05685
K-40	1461	0.80760	0.01487	1.59780	0.02862
Cs-137	662	1.48603	0.02242	2.83317	0.44763

Appendix 5 Listing of gamma spectroscopy standards used for calibration

An up-to-date listing of all gamma spectroscopy standards is maintained in the file 'All-ERISS-standards.xls' on Sharepoint directory :

SSDX > Equipment and Techniques > HPGe Gamma Spec > Calibrations > Standards Listing